

Letter of Notification - Notified body Certificate (CE)

Date: 23.1. 2024

To Whom It May Concern

For the conformity assessment of Risk Class I medical devices supplied in non-sterile condition without measuring function, a legislative exception was used. These Medical Devices do not require the involvement of a Notified Body.

Regards,

Ing. Tomáš Janura, executive

diawin Deutschland GmbH

